

CLEANISEPT® WIPES MAXI

Alcohol-free wipes for rapid disinfection

- Universal, comfortable and time-saving application
- Basic, rapid efficacy
- Especially suited for disinfection of medium-sized surfaces in routine areas



Product description

CLEANISEPT WIPES MAXI are ready to use wipes for rapid disinfection and cleaning of medical devices and medical inventory as well as all kind of surfaces in all areas of the routine. CLEANISEPT WIPES MAXI offer a universal and time-saving application for the disinfection and cleaning of larger surfaces. The solution is alcohol-free and free of aldehydes and phenols. The wipes can be extraxted individually from the re-sealable and user-friendly flowpack.

Range of application

According to Biocidal Products Regulation (BPR): For rapid disinfection and cleaning of alcohol sensitive medical equipment and surfaces of every type.

Acc. to EU Medical Devices Regulation: For rapid disinfection and cleaning of alcohol sensitive, non invasive medical devices, as well as ultrasound probes for abdominal examinations.

Application

Wipe surfaces with CLEANISEPT WIPES MAXI until completely wet and allow to dry for the entire contact time.

Application note

Shelf life after opening: 3 months.

Spectrum of efficacy and contact times			30 s	1 min	2 min			
Bacteria and yeasts								
application recommendation for surface disinfection: Bactericidal, yeasticidal*	VAH**/ EN 16615	with mechanical action, dirty conditions		•				
Enveloped viruses								
active against enveloped viruses	RKI/DVV	incl. HBV/HIV/HCV	•					
	EN 14476	clean conditions	•					
Non-enveloped viruses								
active against papova/SV40 viruses	EN 14476	dirty conditions			•			
active against rotaviruses	RKI/DVV		•					
Additional test results								
bactericidal	EN 1040			•				
	EN 13727	dirty conditions		•				
yeasticidal (C. albicans)	EN 1275			•				
	EN 13624	dirty conditions		•				

^{*} including phase 2 step 1 and phase 2 step 2 tests (quantitative suspension tests and quantitative carrier tests) ** listed in the 5 min column of the VAH list

Composition

100 g solution contain:

0.5 g alkyl(C12-16)dimethylbenzylammoni-umchloride,

0.25 g didecyldimethylammoniumchloride

Manufacturer approval

CLEANISEPT WIPES MAXI are recommended for disinfection of ultrasound units and ultrasound heads of the following manufacturers:

- Aloka GmbH, Meerbusch
- Esaote GmbH, Köln
- Hitachi Medical Systems GmbH, Wiesbaden
- Philips Medizin Systeme GmbH, Hamburg
- Shimadzu Deutschland GmbH, Duisburg
- Siemens Medical Solutions, Erlangen
- SonoAce GmbH, Marl
- Zonare Medical Systems GmbH, Erlangen





listed in the 5 min column of the VAH

CLEANISEPT® WIPES MAXI

Alcohol-free wipes for rapid disinfection

CLEANISEPT WIPES MAXI are recommended for disinfection of medical products of the following manufacturers:

- Cardinal Health, Deutschland
- Karl Storz GmbH & Co. KG, Tuttlingen
- Konrad Hornschuch AG, Weißbach
- Schmitz & Söhne, Wickede
- Tomey/Eye Novation GmbH, Bielefeld

Expert opinions

We have extensive expert reports for all our products. Upon request we are glad to supply you with an expert report portfolio.

Precautionary and hazard statements

Harmful to aquatic life with long lasting effects. Avoid release to the environment. IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Dispose of contents/container to approved disposal company or local collection.

For professional use only.

Manufacturer

Dr.Schumacher GmbH Am Roggenfeld 3, 34323 Malsfeld/Germany Phone: +49 5664 9496-0 Fax: +49 5664 8444 www.schumacher-online.com

Importer

Dr. Schumacher Kimya San.ve Tic. A.Ş İbni Melek Organize Sanayi Bölgesi 5. Yol No:42 Tire-İzmir/Turkey Phone:+90 232 513 5005 Fax:+90 232 853 9487 www.schumacher-online.com.tr

Product	Single unit	Content	Delivery unit	Size	REF
CLEANISEPT WIPES MAXI	pack	10 Wipes	54	20 x 22 cm	00-208-T010EXP
	pack	100 Wipes	6	20 x 22 cm	00-208-T100EXP

The availability of the products and container sizes depend on a successful national registration.

Use disinfectants safely. Always read the label and product information before use.





